



**HAL**  
open science

## High spatial resolution inorganic scintillator detector for high energy X-ray beam at small field irradiation

S.B.C Debnath, C. Fauquet, Agnes Tallet, Anthony Gonçalves, S. Lavandier, F. Jandard, D. Tonneau, Julien Darréon

### ► To cite this version:

S.B.C Debnath, C. Fauquet, Agnes Tallet, Anthony Gonçalves, S. Lavandier, et al.. High spatial resolution inorganic scintillator detector for high energy X-ray beam at small field irradiation. Medical Physics, 2019, 10.1002/MP.14002 . hal-02424913v1

**HAL Id: hal-02424913**

**<https://amu.hal.science/hal-02424913v1>**

Submitted on 28 Dec 2019 (v1), last revised 10 Feb 2020 (v2)

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Copyright

**Title:**

**High spatial resolution inorganic scintillator detector for high energy X-ray beam at small field irradiation**

**Full Authors' names with bylines (in order):**

**Sree Bash Chandra Debnath (S.B.C. DEBNATH<sup>1\*</sup>)**

Aix Marseille Université, CNRS, CINaM UMR 7325, 13288, Marseille, France

**Carole Fauquet (C. FAUQUET<sup>1</sup>)**

Aix Marseille Université, CNRS, CINaM UMR 7325, 13288, Marseille, France

**Agnes Tallet (A. TALLET<sup>2</sup>)**

Institut Paoli-Calmettes, 13009, Marseille, France

**Anthony Goncalves (A. GONCALVES<sup>3</sup>)**

Aix Marseille Université, CNRS UMR 7258, INSERM UMR 1068, CRCM, 13009 Marseille, France

**Sébastien Lavandier (S. LAVANDIER<sup>1</sup>)**

Aix Marseille Université, CNRS, CINaM UMR 7325, 13288, Marseille, France

**Franck Jandard (F. JANDARD<sup>1</sup>)**

Aix Marseille Université, CNRS, CINaM UMR 7325, 13288, Marseille, France

**Didier Tonneau (D. TONNEAU<sup>1</sup>)**

Aix Marseille Université, CNRS, CINaM UMR 7325, 13288, Marseille, France

**Julien Darreon (J. DARREON<sup>2</sup>)**

Institut Paoli-Calmettes, 13009, Marseille, France

**Complete mailing and email address for the Corresponding Author:**

Sree Bash Chandra Debnath

Aix-Marseille University -CINaM-UMR7325

Faculty of Science, Case 913, 13288

Building TPR1, 3<sup>rd</sup> floor, (Room: G.03.23)

Marseille cedex 09, France

**Email:** sree.debnath@univ-amu.fr; sreebashcuet@gmail.com

## High spatial resolution inorganic scintillator detector for high energy X-ray beam at small field irradiation

S.B.C. DEBNATH<sup>1\*</sup>, C. FAUQUET<sup>1</sup>, A. TALLET<sup>2</sup>, A. GONCALVES<sup>3</sup>, S. LAVANDIER<sup>1</sup>,  
F. JANDARD<sup>1</sup>, D. TONNEAU<sup>1</sup>, J. DARREON<sup>2</sup>

<sup>1</sup>Aix Marseille Université, CNRS, CINaM UMR 7325, 13288, Marseille, France

<sup>2</sup>Institut Paoli-Calmettes, 13009, Marseille, France

<sup>3</sup>Aix Marseille Université, CNRS UMR 7258, INSERM UMR 1068, CRCM, 13009  
Marseille, France

### Abstract:

**Purpose:** Small field dosimetry for radiotherapy is one of the major challenges due to the size of most dosimeters, e.g. sufficient spatial resolution, accurate dose distribution and energy dependency of the detector. In this context, the purpose of this research is to develop a small size scintillating detector targeting small field dosimetry and compare its performance with other commercial detectors.

**Method:** An inorganic scintillator detector (ISD) of about 200  $\mu\text{m}$  outer diameter was developed and tested through different small fields dosimetric characterization under high energy photons (6 MV and 15 MV) delivered by an Elekta Linear Accelerator (LINAC). PDD and beam profile measurements were compared using dosimeters from PTW namely, microdiamond and PinPoint 3D detector. A background fiber method has been considered to quantify and eliminate the minimal Cerenkov effect from the total optical signal magnitude. Measurements were performed inside a water phantom under IAEA Technical reports series recommendations (IAEA TRS 381 and TRS 483).

**Results:** Small fields ranging from  $3 \times 3 \text{ cm}^2$ , down to  $0.5 \times 0.5 \text{ cm}^2$  were sequentially measured using the ISD and commercial dosimeters, and a good agreement was obtained among all measurements. The result also shows that, scintillating detector has good repeatability and reproducibility of the output signal with maximum deviation of 0.26% and 0.5% respectively. The Full Width Half Maximum (FWHM) was measured 0.55 cm for the smallest available square size field of  $0.5 \times 0.5 \text{ cm}^2$ , where the discrepancy of 0.05 cm is due to the scattering effects inside the water and convolution effect between field and detector geometries. Percentage Depth Dose (PDD) factor dependence variation with water depth

exhibits nearly the same behavior for all tested detectors. The ISD allows to perform dose measurements at a very high accuracy from low (50 cGy/min) to high dose rates (800 cGy/min) and found to be independent of dose rate variation. The detection system also showed an excellent linearity with dose; hence calibration was easily achieved.

**Conclusions:** The developed detector can be used to accurately measure the delivered dose at small field during the treatment of small volume tumors. The author's measurement shows that despite using a non-water equivalent detector, the detector can be a powerful candidate for beam characterization and quality assurance in e.g., radiosurgery, Intensity Modulated Radiotherapy (IMRT), and brachytherapy. Our detector can provide real-time dose measurement and good spatial resolution with immediate readout, simplicity, flexibility, and robustness.

**Keywords:** *X-ray detector, Inorganic scintillator detector, Small field irradiation, Real-time radiation dose, micro-scintillators.*

## 1. Introduction

Radiation dosimetry plays a very important role in radiotherapy to accurately measure the exact radiation dose delivered to the patients to ensure a high treatment quality assurance [1]. To enhance treatment efficiency, several radiotherapy techniques have been developed for cancer treatment modalities such as Intensity Modulated Radiotherapy (IMRT), Volumetric Modulated Arc Therapy (VMAT), Stereotactic Body Radiation Therapy (SBRT), and Stereotactic Radiosurgery (SRS) etc. These techniques require small irradiation field sizes and high dose spatial gradients to ensure the delivery of accurate high doses with tighter margins around the targeted tumors, enabling possible sparing of organs at risk [2-4]. Unfortunately, accurate measurements are usually hampered due to the size of the conventional dosimeters [5-7] for proper dose distribution in the treatment planning system and patient quality control.

Due to the lack of charged particle equilibrium, chamber size, dose perturbation, corrections of volume averaging effects, and non-equivalence material regarding soft tissue, the measurement with these conventional detectors are complex and hence requires many correction factors specially in the small field dosimetry [7-8]. Hence, several international organizations such as AAPM and IAEA suggested various dosimetry sensors when working

under small fields. Some recent researches indicate that, the suitable detectors for small field dosimetry is plastic scintillation based extradin W1, W2, and radiochromic films, owing to their good correction factor [9-14]. However, the spatial resolution of these detectors is not yet up to the mark due to the minimum size of the sensor head requirement and radiochromic films suffers from time consuming techniques while using. Furthermore, the major drawback of using plastic scintillator based detectors are their high sensitivity to Cerenkov radiation (known as ‘stem’ effect) observed when charge particles generated within the fiber at high energy are slowed down in the fiber core, producing a strong Cerenkov luminescence [15-18].

Hence, we developed an Inorganic Scintillator Detector (ISD) based on a scintillating inorganic cluster optically coupled to a silica optical fiber, very promising on real-time dosimetry [18,20]. Under irradiation, the cluster emits visible light that is driven through the optical fiber toward a photon counter. The detector has been tested for different dosimetric parameters under high energy beams of 5 to 15 MV and experimental conditions close to the real patient treatment scenario. In this context, the aim of this research work is to demonstrate the performances of this ISD detector by quantitative comparison with microdiamond dosimeter (a suitable detector for small fields [8,19]). PinPoint 3D (PP3D), another commercial dosimeter used for regular beam in the patient treatment planning was also considered to show the behavior of ISD with respect to it. Lateral profiles of small and very small size fields (ranging from 3x3 cm<sup>2</sup> down to 0.5x0.5 cm<sup>2</sup>) as well as Percentage Depth Dose (PDD) were systematically performed with all the detectors.

## **2. Materials and Methods**

### **2.1 Source**

Experiments were performed under a LINAC (Elekta Synergy, VERSA) source with the photon beam energy of 6 MV and 15 MV at the Institute Paoli-Calmettes (IPC), Marseille, France. The MLC window of the LINAC can be opened down to a few millimeters length. Elekta LINAC Synergy system can deliver a typical dose rate at 400 MU/min and VERSA can work in two mode of operation (with flattening filter and without flattening filter) in a very high dose rate of 2400 MU/min. The radiotherapy equipment is periodically calibrated so that 1\_MU corresponds to 1\_cGy under reference condition (IAEA TRS 381). The source

can irradiate fields at  $40 \times 40 \text{ cm}^2$ , down to  $0.5 \times 0.5 \text{ cm}^2$  at the iso-center (100 cm from the source). The Elekta system can be rotated up to  $360^\circ$  during irradiation process.

## 2.2 Devices

The novel X-ray probe consists in a 10 m long silica ( $\text{SiO}_2$ ) optical fiber with scintillator clusters grafted at one terminal. Under irradiation, the scintillating clusters fixed at the fiber extremity produce visible light proportional to the irradiation coming from the LINAC source. The visible photons are transmitted through the fiber core toward a photon counter (Aurea<sup>TM</sup>). The core and cladding diameter of the fiber is respectively 100 and 125  $\mu\text{m}$  (Thorlabs<sup>TM</sup>), and the fiber bandwidth is 400-2100 nm. The scintillating material, ZnS:Ag, is a powder of typical 2-3  $\mu\text{m}$  grain size, mixed to PMMA resist diluted in Ethyl lactate ( $\text{C}_5\text{H}_{10}\text{O}_3$ ) solvent. PMMA is a bio-compatible resist commonly used in microelectronics industry. After removing the plastic protective coating from glass optical fiber, the extremity is dipped into this PMMA mixture and immediately removed for a drying step at  $65^\circ\text{C}$ . Consequently a nearly spherical PMMA droplet containing ZnS:Ag clusters is formed. Finally, the device is dipped in a liquid silver paste and dried at room temperature. This latter metallization step leads the device free from any ambient light noise. The head of this detector is shown in compare to other conventional dosimeters (Figure 1b). The sensitive volume is assumed to be a cylinder of diameter 100  $\mu\text{m}$  (fiber core diameter), 1.5  $\mu\text{m}$  lengths. Indeed, light emitted by scintillating grains at distances higher than 1.5  $\mu\text{m}$  from the fiber core is re-absorbed before reaching the core. Thus, we estimate that the sensitive volume of ISD is about  $1.2 \times 10^{-5} \text{ mm}^3$ . This sensitive volume is much lower than that of PinPoint 3D ( $0.016 \text{ cm}^3$  with an outer diameter of 7mm) and microdiamond ( $0.004 \text{ mm}^3$  with an outer diameter of 4.8 mm).

## 2.3 Experimental set-up

Figure 1 represents the detailed experimental set-up used with the LINAC Elekta source and the water phantom. This system is based on a large-size water tank (IBA<sup>TM</sup>) equipped with X, Y, Z stages that allows dose distribution measurements in arbitrary planes, according to international standards such as AAPM TG 142. Results presented in this paper were obtained keeping the beam perpendicular to the water surface and moving the sensor in planes parallel to this surface. The scintillating active part of the fiber based detector is fixed to the scanning unit inside the water tank, whereas the other extremity of the fiber is plugged to a photon

counter The whole set-up is remotely controlled from an external room avoiding any exposure of the electronics to high energy irradiation.

## **2.4 Measurement protocol**

To demonstrate the sensor performance, all the measurements were carried out simultaneously with a microdiamond dosimeter commonly used for small field dosimetry and a Pint Point 3D detector. Otherwise stated, each measurement in this research has been measured inside water phantoms in reference conditions, i.e., SSD (source to water surface distance) of 90 cm and 10 cm depth in water during beam profiling and 100cm SSD during PDD measurements. The sensitive photon counter measure optical signal of the scintillation light in photons per second. Time integration of optical signal gives the total number of photons linked to the irradiation dose. A Matlab<sup>TM</sup> simulator was developed to calculate total amount of photons during each irradiation. In order to compare the performance of scintillator detector with other dosimeters, beam profiles and PDD curves have been normalized.

## **2.5 Cerenkov light subtraction**

The spectral distribution of Cerenkov light is most intense in the blue and ultraviolet regions of the electromagnetic spectrum. This effect has a huge contribution on signal amplitude when using plastic scintillators and optical fibers [15,21-22,26]. Thus, the Cerenkov effect in this case must be considered and removed from the total acquisition signal. In our case, the inorganic scintillator is grafted to a narrow silica fiber core, so that the Cerenkov effect is expected to be weak. Moreover, in order to minimize this effect, the size of the inorganic scintillating head has been reduced as much as possible [21-22]. However, even if the contribution of the Cerenkov Effect is weak, it was systematically quantified and removed from the measurements presented in this article. A background fiber method [15,23-24] has been used to accurately measure the real scintillating light. Indeed, the detector and a bare fiber were simultaneously exposed to radiation at the same x-position (Figure 1a). It relies on the assumption that Cerenkov signal generated in the background fiber is of equal magnitude of the signal fiber. Finally, the actual signal of the detector is obtained by subtracting the signal of the bared fiber from the signal provided by the detector directly read from the photon counter. This Cerenkov correction was considered for all the recorded data given in this study.



### **3. Results**

#### **3.1 Relative dose measurement**

Figure 2a shows the optical signal magnitude variation with time and respective field sizes (ranging from 3 x 3 cm<sup>2</sup> down to 0.5 x 0.5 cm<sup>2</sup>) at 6 MV. These measurements were performed with the sensor placed at the field center in reference conditions. For each field, the acquisition signal was recorded during the delivery of 100 MU dose. Each curve exhibits the same behavior with a rise time of 5 sec followed by a plateau. Finally, a fall time in the microsecond range is observed when the beam is switched off automatically. The photon counter used in this study has a rise and fall time in the ns range (constructor data), so the long raise time demonstrates the LINAC source characteristic. Figure 2a also shows the Cerenkov effect magnitude (blue lines) in the same irradiation conditions. As mentioned in section 2.5, this effect has a very low contribution to the signal recorded by our detector and can be easily eliminated from the total signal. Finally, total number of photons (counts) corresponding to the actual scintillation during irradiation is calculated by integrating the optical signal (free from Cerenkov) with respect to time. This optical signal intensity reported in figure 2b increases with the field size, as the diffusion inside the multi-leaf collimator is increasing. Due to the very low noise level (300 photons/s) and low signal fluctuations (less than 0.2 %) during irradiation signal measured (about 10<sup>5</sup> photons/s), error bar of each ISD measurement is within the dot size on the curve shown in figure 2(b).

#### **3.2 Dose Repeatability and Reproducibility**

Measurement repeatability was checked in standard conditions at 15 MV for 100 MU (Figure 3a) and 20 MU (Figure 3b) doses, delivered at a dose rate of 400 MU/min. For this test, 10 consecutive irradiations were performed on the sensor placed at the center of 1x 1 cm<sup>2</sup> and 0.5 x 0.5 cm<sup>2</sup> fields.

These results show that the ISD demonstrates a very good repeatability with maximum standard deviation of 0.26 % and 1.2% from the average for respective doses of 100 cGy and 20 cGy. The detector was also tested for 7 consecutive days and less than 0.5% day to day

variation of the collected signal was observed. This result demonstrates a very good reproducibility and highlights that ISD does not require regular calibration.

### **3.3 Dose and dose rate linearity**

Detector's linearity was tested from very low dose (5 cGy) to high dose (500 cGy) at the centre of smallest field and observed accurate linear behavior with a linear regression factor of 0.9997, as can be seen in figure 4. Figure 5a shows the variation of the optical signal measured with ISD as a function of time for different dose rates ranging from 50 to 800 cGy/min. The total dose was kept constant at 100 cGy for each experimental measurement point. As an evidence, the higher the dose rate, the shorter is the irradiation period. Figure 5b represents the variation of the total number of photons measured during each irradiation as a function of dose rate. As expected, the detector provides an optical signal that is almost independent of dose rate. Indeed, maximum standard deviation of the signal from the average value is as low as 0.15%, that is a very significant outcome comparing to that obtained with other detectors [see e.g. 14-15,25]. This result shows that ISD can be successfully used at low as well as high dose rate, as it is important for a broad range of radiotherapies.

### **3.4 Beam profiling and comparison**

Beam profiles were measured for fields ranging from  $3 \times 3 \text{ cm}^2$  down to  $0.5 \times 0.5 \text{ cm}^2$  with ISD, PinPoint 3D and microdiamond dosimeters in reference conditions inside water phantoms. We have reported in figures 6(a) to 6(d), the normalized local dose as a function of detector position in crosslink within the field. Step between two successive measurements is maintained at  $200 \mu\text{m}$ . The in-field measurement variation in percentage difference shows only 0% -1.2% with the reference detectors, which is in a very good agreement. The profiles shown in figure 6 exhibit a nearly sharp fall-off at the field edges, as expected for all detectors. Indeed, because of scattered radiation and convolution effect between field and detector geometries [27-29], a perfect fall-off cannot be achieved. Both effects are usually gathered in the penumbra region. Convolution effect contribution is roughly proportional to the detector sensitive head size. In our case, thanks to the small size of ISD, the penumbra profile (between 20-80%) is sharper in comparing to the other dosimeters considered in this research.

Using ISD, the measured Full Width Half Maximum (FWHM) are 3, 2, 1 and 0.55 cm, in a very good agreement with selected beam size. Little discrepancies between ISD and the microdiamond detector at field edges is observed at a level around 4.45% to 13.3 %.

### **3.5 PDD measurements and comparison**

The central axis dose distribution is one of the most important clinical parameters and it is typically characterized by Percentage Depth Dose factor (PDD) measurements. The PDD curves obtained with the ISD is presented in figure 7 for different small fields varying from 3 x 3 cm<sup>2</sup> down to 0.5 x 0.5 cm<sup>2</sup> and compared to that measured with microdiamond. Due to the increasing number of scattered photons and secondary electrons created by the incident high energy beam penetration, the PDD curve steeply increases till a maximum depth dose in the build-up region. After the maximum value, the PDD smoothly decreases with the depth due to absorption of generated photons and/or charged particles by water. With ISD, the maximum PDD value is found at 15 mm from the water surface, a standard and expected value for the 6 MV incident photon beam [15,27-28,30]. This maximum dose depth is independent of the field size considered here. A very good agreement is found between the measurements given by the three detectors for 3 x 3 cm<sup>2</sup> and 2 x 2 cm<sup>2</sup> fields. However, far from the maximum dose depth, differences between detectors become visible and are more pronounced at small fields.

The PDD difference between ISD and microdiamond detector varies from 0.1% up to 13.3%, whereas the average percentage difference stays in a reasonable range, from 2.1% to 5.6 % for all the measurements. Note that, the PDD difference around the maximum of the build-up region stays below 1% for all the small fields.

## **4. Discussion**

The developed ISD sensor detects a maximum dose delivery at a depth of approximately 15 mm inside water, value in perfect agreement with commonly known. This value for maximum dose is independent of the field size from 3 x 3 cm<sup>2</sup> down to 0.5 x 0.5 cm<sup>2</sup>. The PDD exhibits almost the same behavior for all detectors unless a little discrepancy observed for ISD, especially in the water depths larger than 15 mm. This effect can be strongly reduced by necessary corrections of the reference detector [31-33]. Moreover, some discrepancies that

appears in small fields and nearby water surface can be due to the proper centering of the reference dosimeter. However, discrepancies observed at higher depth can be attributed to the different sensitivities of each detector to incident and secondary photons as well as other charged particles.

The ISD detector has a very small scintillating active part ( $1.2 \times 10^{-5} \text{ mm}^3$ ) and thus allows to measure more accurately dose distribution at the edge of the beam profile. Indeed, the minimum step size of 200  $\mu\text{m}$  available in the scanning unit (IBA<sup>TM</sup>) was selected to characterize all the beam profiles.

For the  $0.5 \times 0.5 \text{ cm}^2$  field, we measured an experimental FWHM of 0.55 cm. The small discrepancy between expected and measured field sizes can be explained by three different effects. First, the measured field is wider due to scattered photons on the edge of the thick lead collimator of the LINAC. Secondly, the field also expanded due to charged particle generation inside water at the detector surface vicinity. Finally, there must be considered the convolution effect between the actual field geometry (square of 5 mm side) and the detector geometry. This latter effect leads to an apparent side increase of 100  $\mu\text{m}$ , that corresponds to the ISD sensor diameter.

The minimum step available on PTW translation stage is 200  $\mu\text{m}$ . In these conditions, the signal difference between two consecutive measurements in the penumbra region is about 4000 photons/s, which is much higher than the photon counter sensitivity (20 photons/s). Thus, the step size could be decreased down to 100  $\mu\text{m}$  and even less using a lower diameter core fiber keeping a significant signal to noise ratio. Note that, we already presented a sub 3  $\mu\text{m}$  lateral resolution ISD, designed for low energy (2 to 30 keV) X-ray beam profiling [34]. In this study, we found a detection flux threshold as low as  $10^3$  X-ray photons/s/ $\mu\text{m}^2$ . Thus, the design of an ISD detector for high energy beam offering a spatial resolution much better than 100  $\mu\text{m}$  is realistic.

In contrast to plastic scintillating detector (PSD), Cerenkov contribution in the developed detector was found to be negligible regarding the magnitude of the optical signal measured in the experimental conditions. Moreover, the ambient optical noise coming from the experimental room was avoided by coating the sensor head with a very thin metallic layer embedding the scintillators. Thus, scintillating detector in this regard exhibits higher signal to noise ratio.

Device stability and reproducibility tested at various dose rates ranging from 50 to 800 cGy/min will allow to use it at low dose application to the small size tumor and high dose delivery in the brachytherapy or radiosurgery. ISD is based on optical transitions that are less sensitive to ambient temperature and pressure variations than electrical charge-based detectors. Therefore, it is anticipated that the ISD detector is free from any perturbations due to temperature and pressure variations. This result is still under investigation by relevant experimentation.

Finally, the ISD provides  $4.75 \times 10^5$  visible photons collected with the photon counter for a total dose of 100 cGy, independently of the dose rate (Figure 5). The calibration is thus about 4700 visible photons generated per cGy of dose delivered. As the detector's signal magnitude is proportional to the dose, this system can easily be used on patients for radiotherapy dose measurement for small field irradiation with better accuracy.

## **5. Conclusion**

Due to the tiny scintillating sensitive volume used at the optical fiber extremity (about  $1.2 \times 10^{-5} \text{ mm}^3$ ), the developed detector in this article exhibits a very high lateral resolution in dosimetry measurements. Indeed, it allows to accurately define the lateral profiles of very small fields down to  $0.5 \times 0.5 \text{ cm}^2$  with a precision of around  $200 \text{ }\mu\text{m}$ . Owing to the detector's minimum size (nearly an ideal point detector), ISD is almost free from detector edge effects and no aberrant measurements were observed, while it was the case for the concurrent microdiamond and PinPoint 3D detectors. The high spatial resolution of the ISD was examined during the sharp fall-off in the penumbra region of the smallest beam profiling in compare to the commercial detector.

The real-time inorganic detector in this study is fast, robust, non-sensitive to external noise and stem effect. As expected, the integrated total relative dose measurement in this study was found to be independent of the dose rate variability. Moreover, a perfect linearity with dose was observed for the smallest field considered in this study and lead to detector calibration of  $4.7 \times 10^5$  photons collected/Gy .

Further analysis on detector characterization and resolution improvement is ongoing, whereas the results presented in this research demonstrate the prospects of ZnS:Ag based scintillator detector. The performance of ISD shown in this study demonstrates that the detector can be a suitable candidate for small field dosimetry, a proper quality control tool for possible early stage tumor treatment. However, further investigation of the detector will follow the

comparison with water equivalent scintillating dosimeters e.g. PSD, radiochromic films etc. to make a proper comparison during absolute dose measurements.

## 6. Acknowledgements

This project has received funding from the European Union's Horizon 2020 Research and Innovation Program under the Marie Skłodowska-Curie grant agreement No.713750. Also, it has been carried out with the financial support of the Regional Council of Provence- Alpes-Côte d'Azur and with the financial support of the A\*MIDEX (n° ANR- 11-IDEX-0001-02), funded by the Investissements d'Avenir project funded by the French Government, managed by the French National Research Agency (ANR)

## 7. Conflict of Interest

The authors have no conflict of interest to disclose.

## 8. References

1. Vordermark D. Ten years of progress in radiation oncology. *BMC Cancer*. 2011;11(1):503. doi:10.1186/1471-2407-11-503
2. Arnfield MR, Wu Q, Tong S, Mohan R. Dosimetric validation for multileaf collimator-based intensity-modulated radiotherapy: A review. *Med Dosim*. 2001;26(2):179-188. doi:10.1016/S0958-3947(01)00058-9
3. Duggan DM, Coffey CW. Small photon field dosimetry for stereotactic radiosurgery. *Med Dosim*. 1998;23(3):153-159. doi:10.1016/S0958-3947(98)00013-2
4. Galvin JM, Ezzell G, Eisbrauch A, et al. Implementing IMRT in clinical practice: A joint document of the American Society for Therapeutic Radiology and Oncology and the American Association of Physicists in Medicine. *Int J Radiat Oncol Biol Phys*. 2004;58(5):1616-1634. doi:10.1016/j.ijrobp.2003.12.008
5. Stasi M, Baiotto B, Barboni G, Scielzo G. The behavior of several microionization chambers in small intensity modulated radiotherapy fields. *Med Phys*. 2004;31(10):2792-2795. doi:10.1118/1.1788911
6. Lauba WU, Wong T. The volume effect of detectors in the dosimetry of small fields used in IMRT. *Med Phys*. 2003;30(3):341-347. doi:10.1118/1.1544678
7. Parwaie, Wrya, et al. "Different dosimeters/detectors used in small-field dosimetry: Pros and cons." *Journal of medical signals and sensors* 8.3 (2018): 195.
8. Palmans, H., et al. "Dosimetry of small static fields used in external beam radiotherapy: an IAEA-AAPM international code of practice for reference and relative dose determination." Vienna: International Atomic Energy Agency (2017)
9. Huet, C., et al. "Characterization and optimization of EBT2 radiochromic films dosimetry system for precise measurements of output factors in small fields used in radiotherapy." *Radiation Measurements* 47.1 (2012): 40-49.
10. Therriault-Proulx, Francois, et al. "Quality assurance for gamma knife perfexion using the exradin W1 plastic scintillation detector and Lucy phantom." *Physics in Medicine & Biology* (2019).
11. Galavis, Paulina E., et al. "Characterization of the plastic scintillation detector Exradin W2 for small field dosimetry." *Medical physics* 46.5 (2019): 2468-2476.

12. Hoehr, C., et al. "Characterization of the exradin W1 plastic scintillation detector for small field applications in proton therapy." *Physics in Medicine & Biology* 63.9 (2018): 095016.
13. Gonzalez-Lopez, Antonio, Juan-Antonio Vera-Sanchez, and Jose-Domingo Lago-Martin. "Small fields measurements with radiochromic films." *Journal of Medical Physics/Association of Medical Physicists of India* 40.2 (2015): 61.
14. Carrasco P, Jornet N, Jordi O, et al. Characterization of the Exradin W1 scintillator for use in radiotherapy. *Med Phys*. 2015;42(1):297-304. doi:10.1118/1.4903757
15. Beddar AS, Mackie TR, Attix FH. Water-equivalent plastic scintillation detectors for high-energy beam dosimetry: I. Physical characteristics and theoretical considerations. *Phys Med Biol*. 1992;37(10):1883-1900. doi:10.1088/0031-9155/37/10/006
16. Beddar AS, Mackie TR, Attix FH. Water-equivalent plastic scintillation detectors for high-energy beam dosimetry. *Phys Med Biol*. 1992;37(10):1883-1900. <http://iopscience.iop.org/0031-9155/37/10/006/pdf/pb921006.pdf>.
17. Gagnon JC, Thériault D, Guillot M, et al. Dosimetric performance and array assessment of plastic scintillation detectors for stereotactic radiosurgery quality assurance. *Med Phys*. 2012;39(1):429-436. doi:10.1118/1.3666765
18. Andersen, Claus E. "Fiber-coupled Luminescence Dosimetry in Therapeutic and Diagnostic Radiology." AIP Conference Proceedings. Vol. 1345. No. 1. AIP, 2011.
19. De Coste, Vanessa, et al. "Is the PTW 60019 microDiamond a suitable candidate for small field reference dosimetry?." *Physics in Medicine & Biology* 62.17 (2017): 7036.
20. Kertzscher, G., and Sam Beddar. "Inorganic scintillator detectors for real-time verification during brachytherapy." *Journal of Physics: Conference Series*. Vol. 847. No. 1. IOP Publishing, 2017.
21. Beaulieu L, Beddar S. Review of plastic and liquid scintillation dosimetry for photon, electron, and proton therapy. *Phys Med Biol*. 2016;61(20):R305-R343. doi:10.1088/0031-9155/61/20/R305
22. Ramírez M, Martínez N, Marcazzó J, Molina P, Feld D, Santiago M. Performance of ZnSe(Te) as fiberoptic dosimetry detector. *Appl Radiat Isot*. 2016;116:1-7. doi:10.1016/j.apradiso.2016.07.007
23. Létourneau D, Pouliot J, Roy R. Miniature scintillating detector for small field radiation therapy. *Med Phys*. 1999;26(12):2555-2561. doi:10.1118/1.598793
24. Liu PZY, Suchowerska N, Lambert J, Abolfathi P, McKenzie DR. Plastic scintillation dosimetry: Comparison of three solutions for the Cerenkov challenge. *Phys Med Biol*. 2011;56(18):5805-5821. doi:10.1088/0031-9155/56/18/003
25. Alharbi M, Gillespie S, Woulfe P, McCavana P, O'Keeffe S, Foley M. Dosimetric Characterization of an Inorganic Optical Fiber Sensor for External Beam Radiation Therapy. *IEEE Sens J*. 2019;19(6):2140-2147. doi:10.1109/JSEN.2018.2885409
26. Linares Rosales, Haydee M., et al. "Optimization of a multipoint plastic scintillator dosimeter for high dose rate brachytherapy." *Medical physics* 46.5 (2019): 2412-2421.
27. Yoo WJ, Moon J, Jang KW, et al. Integral T-shaped phantom-dosimeter system to measure transverse and longitudinal dose distributions simultaneously for stereotactic radiosurgery dosimetry. *Sensors (Switzerland)*. 2012;12(5):6404-6414. doi:10.3390/s120506404
28. Lee B, Jang KW, Cho DH, et al. Measurement of Two-Dimensional Photon Beam Distributions Using a Fiber-Optic Radiation Sensor for Small Field Radiation Therapy. *IEEE Trans Nucl Sci*. 2008;55(5):2632-2636. doi:10.1109/TNS.2008.2002579
29. Petric MP, Robar JL, Clark BG. Development and characterization of a tissue equivalent plastic scintillator based dosimetry system. *Med Phys*. 2006;33(1):96-105. doi:10.1118/1.2140118
30. Fontbonne JM, Iltis G, Ban G, et al. Scintillating fiber dosimeter for radiation therapy accelerator. *IEEE Trans Nucl Sci*. 2002;49

31. Das, I., Y. Akino, and P. Francescon. "TU-F-BRE-05: Experimental Determination of K Factor in Small Field Dosimetry." *Medical Physics* 41.6Part27 (2014): 469-469.
32. Das, Indra, and Paolo Francescon. "Comments on the TRS-483 protocol on small field dosimetry." *Medical physics* 45.12 (2018): 5666-5668.
33. Underwood, T. S. A., et al. "Application of the Exradin W1 scintillator to determine Ediode 60017 and microDiamond 60019 correction factors for relative dosimetry within small MV and FFF fields." *Physics in Medicine & Biology* 60.17 (2015): 6669.
34. Xie Z, Maradj H, Suarez M-A, et al. Ultracompact x-ray dosimeter based on scintillators coupled to a nano-optical antenna. *Opt Lett.* 2017;42(7):1361. doi:10.1364/ol.42.001361

## Figure Legends:

### Figure 1:

- (a) Elekta LINAC source equipped with a PTW<sup>TM</sup> motorized 3D water phantom.
- (b) ISD head dimension (below) compared to microdiamond (middle) and PinPoint 3D (top) dosimeters.

### Figure 2:

- (a) ISD output signal with time for different fields of 100 MU doses delivered at 400 MU/min
- (b) Integrated output signal as a function of field size.

### Figure 3:

- (a) Repeatability of measurement with ISD for ten successive irradiations at 100 MU.
- (b) Repeatability of measurement with ISD for ten successive irradiations at 20 MU.

### Figure 4: ISD Signal Linearity with Dose.

### Figure 5:



- (a) Variation of signal as a function of time for 100 cGy delivered with dose rates ranging from 50 to 800 cGy/min.
- (b) Integrated number of visible photons generated by the ISD corresponding to figure 5(a).

**Figure 6:** Lateral beam profiles obtained using ISD (red), Microdiamond (blue line) and PinPoint 3D (black dashed line) for 6(a) 3x3 cm<sup>2</sup>, 6(b) 2x2 cm<sup>2</sup>, 6(c) 1x1 cm<sup>2</sup> and 6(d) 0.5x0.5 cm<sup>2</sup> field sizes. The brown dashed line shows the percentage difference in the measurements between ISD and




microdiamond.


**Figure 7:** PDD measurements obtained using ISD (red) compared to the microdiamond (dashed dark blue line) and PinPoint 3D (black) for 7(a)  $3 \times 3 \text{ cm}^2$ , 7(b)  $2 \times 2 \text{ cm}^2$ , 7(c)  $1 \times 1 \text{ cm}^2$  and 7(d)  $0.5 \times 0.5 \text{ cm}^2$  fields. The brown dashed line shows the difference between measurements given by ISD and microdiamond.




mp\_14002\_f1.jpg




mp\_14002\_f2.jpg




mp\_14002\_f3.jpg




mp\_14002\_f4.jpg



mp\_14002\_f5.jpg



mp\_14002\_f6.jpg



mp\_14002\_f7.jpg